

Note to document the ReTest of the A-G 81BF1216-G Pressure Relief Valve on the 100 L LN2 Dewar at Lab B

November 3, 2015

H. T. Diehl, A. Lathrop

We use a 100 L Dewar outside of Lab B as part of a test stand for the DECam LN2 pumps. This system is documented and the documentation is stored in <http://des-docdb.fnal.gov:8080/cgi-bin/ShowDocument?docid=3678>. SV6213, an Anderson-Greenwood (AG)81BF1216-G 150 psig, which is the main pressure relief valve, was removed for recertification. It was retested (passed) and retagged in the PPD Calibration Shop by Mr. Barger. It was returned to us today. It will be returned to the test stand for operation, which will commence as soon as is practical.

This note will be added to docdb 3678.